



IHE Integration Statement		Date: Jun. 21 th , 2022
Vendor	Product Name	SW Build
ESAOTE S.p.A.	MyLab 9, X5, X6/X7, X8, X9, X75, X90, Omega and Sigma Platforms; MyLab XPro30, MyLab XPro80 * <i>* all with DICOM[®] option</i>	F11XXXX
These products are intended to implement all transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors and Options listed below:		
Integration Profiles Implemented	Actors Implemented	Options Implemented
Scheduled Workflow	Acquisition Modality	Broad Worklist Query
Patient Information Reconciliation	Acquisition Modality	none
Portable Data for Imaging	Portable Media Creator	none
Evidence Documents	Acquisition Modality	none
Echocardiography Workflow	Acquisition Modality	Broad Worklist Query Stress Echo
Cardiology Evidence Documents	Acquisition Modality	Echo Evidence
Consistent Time	Time Client	none
Internet address for vendor's IHE information: http://www.esaote.com		
Vendor's e-mail address for information about IHE: ihe@esaote.com		
Links to Standards Conformance Statements for the Implementation		
DICOM[®]	http://www.esaote.com/dicom.htm	
Links to general information on IHE		
In North America: www.ihe.net	In Europe: www.ihe-europe.net	In Japan: www.ihe-j.org

ESAOTE S.p.A.
via Melen 77
16152 Genova (Italy)

© Copyright ESAOTE S.p.A. 2022. All rights reserved.

* DICOM is the registered trademark of the National Electrical Manufacturers Association for its standards publications relating to digital communications of medical information.